



Shenzhen Sinotest Measurement Technology Co.,Ltd.



### Calibration Certificate

1 3

Page of

Certificate No. ST202409007219

Client Name \_\_\_\_\_

Address 14718 #6 #10 \$

Description ' ( )VOCs \* + ,

Model/Type FID3 Manufacturer \_\_\_\_\_

Serial No. 3081017 Management No. 6 7 /

Date of Receipt 2024 < Year 08 = Month 31 : Day

Calibration > ? : 08 = Month 31 : Day

08 = Month 30 : Day

08 = Year 08 = Month 31 : Day



Issued by D 23 ( F G H )

Approved by 李星

Inspected by 郑晓萍

Calibrated by 李星

NO P Q R S T U V W X Y Q Z [ \ 4 ] ^ \_ ` a b \_ c d  
B e 3 \$ f 4 \$

ADD N Floor 3, floor 4, building B, Haoye branch, Baoan Taohuayuan  
Science and Technology Innovation Park, Fenghuanggang community,  
Xixiang Street, Bao 'an district, Shenzhen

g N518102  
h i N0755-27780080  
j Nwww.sinotest.cn  
Email N888@sinotest.cn





### Results of Calibration

Certificate No. ST202409007219

3 3  
Page of

1. k l m n o p \* q N  
(Appearance & Working Performance Check) r s(Pass)

2. t u v w N

Indication Error

x ? u Standard Value	t u Indication	v w Error	y W MPE	Z { Pass/Fail
500  mol/mol	501  mol/mol	+0.2%	±10%	P
2003  mol/mol	1999  mol/mol	-0.2%	±10%	P
10002  mol/mol	10000  mol/mol	0.0%	±10%	P

3. } ~ o N

Repeatability

x ? u Standard Value	} ~ o Repeatability	] • € • Technical Requirement	Z { Pass/Fail
500  mol/mol	0.6%	,2%	P

4. f „ ... † N

Response Time

x ? u Standard Value	f „ ... † Response Time	] • € • Technical Requirement	Z { Pass/Fail
10002  mol/mol	4.00s	,10s	P

† ^: % C + Š z ‹ œ • Ž • • ‘ ‘ “ N  
(The Expanded Uncertainty of the Measurement Result Is)

$U_{rel}=1.5\%$ ; ... † N=0.5s k=2

( “ • JJF1059.1-2012 + Š • • ‘ ‘ – ‘ ~ † TM  
(According to JJF1059.1-2012 Evaluation and Expression of Uncertainty in Measurement)

----- § B ‹ œ(Blank below)-----









中国认可  
国际互认  
校准  
CALIBRATION  
CAPABILITY

# Calibration Certificate

1 3

Certificate No. ST202409007198

Page of

Client Name \_\_\_\_\_

Address 4718 6 10

Description VOCs

Model/Type FID4 Manufacturer \_\_\_\_\_

Serial No. 1021009 Management No. /

Date of Receipt 2024 Year 08 Month 31 Day

\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_



DIRECTIONS

CertificateNo.

Page of

/ fi fl l i Ž ž ! . " # \$ % & ' ( ) \* + ,

- fi . ~ / ! 0 1 2 3 4 5 6 7 8 9 : ; < = 9 : > ? @ A B C \$ % D ' ,

" fi . & ' ~ E F ! 0 1 > G H I J . K L & ' ! \$ % M N O P ,

Q fi . & ' ~ N R S / T U f l l i j ~ V W O P ,

X fi Y Z . B l [ ' U \ ] ^ \_ ` a f i b C ? c d ~ ! e f , ` a g ~ Y h i f l l ~ j C k l m ,

n fi ~ o ! p 3 D q r a # s R m U t u v w x y z h i ! p 3 D q s l m D q ,

{ fi fl l j . ~ ! e f | R } ~ € , f ,

CNAS

...fi . K & ' ! t † ^ 1 % W S fl l < € ' p 3 ' " U " u ' " ! 0 1 \ • p 3 ' " e -

™ S z > œ T R \* Ÿ j ç é æ ¥ < | § " © ª « = & ' ñ '

— ®

R m ± 6

4 5 \* +

² ³ / ´ μ ³ ¶ D ' Ÿ .

, ¹ º » ! ³ ¶ ~

¼ R ½ B l

¾ ª ç Å

À Á Â \$ % » Ä Å

Æ Ç

& ' É É

É É # v l í o í ° í ¹ Ð Ñ Ò Ó Ñ Ò Ó × Ø Ð Ù ° Ù

Ò Ú Ý W

Þ " " " " ß ~ ~ ~ ~ à á J á ß

á ä å Ç à æ

ç è & ' é ê ä Ç



Shenzhen Sinotest Measurement Technology Co.,Ltd.

## Results of Calibration

Certificate No. ST202409007198

3  
Page of 3

1. klmnop  
(Appearance & Working Performance Check)

rs





## DIRECTIONS

CertificateNo. \_\_\_\_\_

Page \_\_\_\_\_ of \_\_\_\_\_

/ fi fl Ł Ł Ź Ź ! . " # \$ % & ' ( ) \* + ,

- fi . ~ / ! 0 1 2 3 4 5 6 7 8 9 : ; < = 9 : > ? @ A B C \$ % D ' ,

" fi . & ' ~ E F ! 0 1 > G H I J . K L & ' ! \$ % M N O P ,

Q fi . & ' ~ N R S / T U fi Ł Ł I J ~ V W O P ,

X fi Y Z . B Ł [ ' U \ ] ^ \_ ` a fi b C ? c d ~ ! e f , ` a g ~ Y hi fi Ł Ł ~ j C k l m ,

n fi ~ o ! p 3 D q r a # s R m U t u v w x y z hi ! p 3 D q s l m D q ,



### Results of Calibration

CertificateNo. ST202409007238

3 Page 3 of

1. klmnop  
(Appearance & Working Performance Check) rs(Pass)

2. tuvwn

Indication Error

x <sub>0</sub> StandardValue	t <sub>u</sub> Indication	v <sub>w</sub> Error	MPE	z Pass/Fail
500μol /mol	501μol /mol	+0. 2%	± 10%	P
2003μol /mol	2005μol /mol	+0. 1%	± 10%	P
10002μol /mol	9999μol /mol	0. 0%	± 10%	P

3. φN

Repeatability

x <sub>0</sub> StandardValue	φ Repeatability	ε TechnicalRequirement	z Pass/Fail
500μol /mol	0. 8%	2%	P

4. η.

Response Time

x <sub>0</sub> StandardValue	η. ResponseTime	ε TechnicalRequirement	z Pass/Fail
10002μol /mol	4. 03s	10s	P

100%SE

(The Expanded Uncertainty of the Measurement Result Is)

$$U_{rel}=1.5\%; N. U=0.5s k=2$$

(JJF1059.1-2012+GTM)

(According to JJF1059.1-2012 Evaluation and Expression of Uncertainty in Measurement)

-----Blank below)-----

|





## Results of Calibration

CertificateNo. ST202409007275

3 3  
Page of

---

1. klmnop  
(Appearance & Working Performance Check)

rs(Pass)

---

2. tuvwn  
Indication Error

---

# 校准证书

## Calibration Certificate

中国认可  
CALIBRATION

Z2024N2-H195899

Page of

客户名称

**Client Name**

上海汉洁环境工程有限公司

地址

**Address**

上海市静安区共和新路4718弄6号楼10层

GM8910

BENETECH

# 说 明

## DIRECTIONS

证书编号 Z2024N2-H195899  
*Certificate No.*

第 2 页  
*Page*

十位数字后跟四位小数为 ISO 15001 2015 标准

[The remainder of the page is obscured by heavy black redaction bars.]

# 校准结果

## Results of Calibration



证书编号  
Certificate No. Z2024N2-H195899

第 3 页 共 页  
Page of

1. 外观及工作性能检查:  
(Appearance & Working Performance Check)

符合 (Pass)

2. 风速校准:  
Wind Speed Calibration

风速点 Wind Point	标准风速 Standard Wind Speed Value	风速示值 Wind Speed Indication	修正值 Correction Value	 MPE	结论 Pass/Fail
2m/s	2.00m/s	2.0m/s	0.0m/s	±0.60m/s	P
5m/s	5.00m/s	5.2m/s	-0.2m/s	±0.75m/s	P
10m/s	10.00m/s	10.3m/s	-0.3m/s	±1.00m/s	P
15m/s	15.00m/s	15.5m/s	-0.5m/s	±1.25m/s	P
20m/s	20.00m/s	20.6m/s	-0.6m/s	±1.50m/s	P
25m/s	25.00m/s	25.6m/s	-0.6m/s	±1.75m/s	P
30m/s	30.00m/s	30.7m/s	-0.7m/s	±2.00m/s	P

说明:本次测量结果的扩展不确定度为:

(The Expanded Uncertainty of the Measurement Result Is)

(2~5)m/s,  $U=0.20$ m/s; (>5~30)m/s,  $U=0.60$ m/s  $k=2$

(依据JJF1059.1-2012测量不确定度评定与表示)

地 址

上海市静安区共和新路4718弄6号楼10层

Address

仪器名称

CM8910  
风速仪 (温湿度部分)

Description

Manufacturer

BENETECH

出厂编号

2599801

Serial Number

管理编号

Management No.

08

Year

Month

校准日期

2024 年

08 月

16 日

Calibration Date

Year

Month

Day

Due Date

Day

发布日期

Issue Date

Month

Day

发证单位(专用章)

Issued by (stamp)

批准:

Approved by

核验:

Inspected by

校准:

Calibrated by

地址: 广东省深圳市龙岗区锦龙大道2号

邮编 (Post Code): 518116

# 说 明

## DIRECTIONS

证书编号 Z20241-H211130  
*Certificate No.*

第 2 页 共 3 页  
*Page of*

1. 本实验室质量管理体系依据ISO/IEC17025:2017建立。  
The laboratory quality management systems document is established according to ISO/IEC17025:2017.





证书编号 Z20241-H211130  
Certificate No.

第 3 页 共 3 页  
Page of

1. 外观及工作性能检查:  
(Appearance & Working Performance Check)

符合 (Pass)

2. 湿度修正值 (T=20°C)

Humidity correction value

校准点 Calibration Point	标准值 Standard Value	示值 Indication	修正值 Correction	允差 MPE	结论 Pass/Fail
40%RH	40.1%RH	44.3%RH	-4.2%RH	±5.0%RH	P

[Redacted]
[Redacted]

校准点 Calibration Point	标准值 Standard Value	示值 Indication	修正值 Correction	允差 MPE	结论 Pass/Fail
--------------------------	-----------------------	------------------	-------------------	-----------	-----------------

[Redacted]
------------

说明:本次测量结果的扩展不确定度为:  
(The Expanded Uncertainty of the Measurement Result Is)

湿度:  $U=1.5\%RH$  温度:  $U=0.3^{\circ}C$   $k=2$

[Redacted]
[Redacted]

# 校准证书

Calibration Certificate

中国认可  
国际互认  
校准

证书编号  
Certificate No. 720242-1105992

Page

客户名称  
Client Name 上海汉洁环境工程有限公司

地址  
Address 上海市静安区共和新路4718弄6号楼10层

仪器名称  
Description 风速仪（大气压部分）

型号/规格  
Model/Type GM8910 制造厂商  
Manufacturer BENETECH

出厂编号  
Serial Number 2599801 管理编号  
Management No.

接收日期  
Date of Receipt 2024 年 08 月 15 日  
Year Month Day

校准日期  
Calibration Date 2024 年 08 月 17 日  
Year Month Day

建议下次校准日期  
2025 年 08 月 16 日  
Month Day

发布日期  
Issue Date 2024 年 08 月 17 日  
Year Month Day

发证单位(专用章)  
Issued by (stamp)

批准:  
Approved by

(科室主任)

核验:  
checked by

李娟

校准:  
Calibrated by

地址: 广东省深圳市龙岗区锦龙大道2号  
ADD: No.2, Jinlong Avenue, Longgang District, Shenzhen, Guangdong, China  
电话 (TEL): 0755-84815081

邮编 (Post Code): 518116  
网址: <http://www.tiansu.org>  
Email: [zskf@tiansu.org](mailto:zskf@tiansu.org)



明

DIRECTIONS

证书编号 Z20242-H195992  
*Certificate No.*

第 2 页 共 页  
*Page*

本实验室质量管理体系文件按照ISO/IEC17025:2017建立

The laboratory quality management systems document is established according to ISO/IEC17025:2017.

